

## Abstract

# Pilot Development of BP-Glass for Unobtrusive Ambulatory Blood Pressure Monitoring

Quan Zhang<sup>1</sup>, PhD; Gang Hu<sup>1</sup>, PhD; Xiangguo Yan<sup>2</sup>, PhD; Kong Yien Chin<sup>3</sup>, PhD; Gary E Strangman<sup>1</sup>, PhD; Ni Zhao<sup>4</sup>, PhD; Yuan-Ting Zhang<sup>5</sup>, PhD

<sup>1</sup>Neural Systems Group, Department of Psychiatry, Harvard Medical School and Massachusetts General Hospital, Charlestown, MA, United States

<sup>2</sup>Key Laboratory of Biomedical Information Engineering of Education Ministry, Xi'an Jiaotong University, Xi'an, China

<sup>3</sup>Medical Physics Dept, Leicester Royal Infirmary, University Hospitals of Leicester, Leicester, United Kingdom

<sup>4</sup>Department of Electronic Engineering, Hong Kong, China (Hong Kong)

<sup>5</sup>Department of Electronic Engineering, The Chinese University of Hong Kong, Hong Kong, China (Hong Kong)

### Corresponding Author:

Quan Zhang, PhD

Neural Systems Group

Department of Psychiatry

Harvard Medical School and Massachusetts General Hospital

13th St, Building 149, Rm 10029

Charlestown, MA, 02129

United States

Phone: 1 617 724 9608

Fax: 1 617 726 4078

Email: [qzhang@nmr.mgh.harvard.edu](mailto:qzhang@nmr.mgh.harvard.edu)

## Abstract

**Background:** Cardiovascular disease (CVD) is the leading cause of death and lower life expectancy worldwide, and is a heavy financial burden both to individuals and to the health system. Blood pressure (BP) control is one of the key strategies to prevent CVDs such as heart attacks and strokes; unfortunately, the proportion of patients with controlled BP has not improved in recent years, despite the increased intensity of therapy. Efforts are being made to improve this situation, including early identification of at-risk individuals using connected health concepts, individualized therapies, adoption of more sensitive clinical indices, and home BP monitoring. The existing BP methods each demonstrate different limitations, including their obtrusiveness, inability to provide continuous monitoring, difficulty integrating into Electronic Health Records, human error concerns, or high cost. New technologies for unobtrusive ambulatory blood pressure monitoring (ABPM) remain to be developed to combine the latest sensor and information technologies and to support the coming connected health strategies.

**Objective:** The objective of this pilot study is to develop a novel, unobtrusive ABPM technology that features beat-to-beat continuous BP monitoring, seamless automatic integration with electronic health records (EHR) and/or internet, low cost to meet today's need of BP control and CVD prevention.

**Methods:** In this paper, we describe the pilot development of a novel wearable ABP monitoring device, BP-Glass. BP-Glass integrates three noninvasive, cuffless and continuous BP sensing technologies into a pair of eyeglasses: tonometry, pulse transit time (PTT) and hemodynamics-based technologies. In the BP-Glass prototype, a superficial temporal artery (STA) tonometry sensor is integrated into a pair of eyeglasses for unobtrusive sensing of ABP; together with a battery powered wearable recording system, the BP-Glass system is capable of recording ABP, electrocardiogram (ECG), respiration, cerebral hemodynamics, systemic hemodynamics and subjects' motion for up to 24 hours at 250Hz sampling rate, and the total weight of the whole multi-modality monitoring system is less than 350 grams. The system's performance is demonstrated by ambulatory tests during subject's normal activity; ABP fluctuations during micturition and Valsalva maneuvers are discussed in detail.

**Results:** Successful ambulatory BP monitoring using BP-Glass during people's daily activities were conducted. Interesting spontaneous or introduced blood pressure fluctuations were captured during the ambulatory recording, including Micturition response and Valsalva Maneuver response.

**Conclusions:** These preliminary results demonstrate the feasibility of our BP-Glass design for ABP monitoring, and suggest that system has significant potential in the diagnosis and management of cardiovascular and cerebrovascular disease, especially

powerful in catching ABP fluctuations during transient symptoms, such as syncope, with unpredictable onsets. To the best of our knowledge, this is the first report of multi-modality ABP monitoring and simultaneous blood pressure and hemodynamics recording during events in peoples' daily activity.

(*iProc 2015;1(1):e8*) doi: [10.2196/ipro.4691](https://doi.org/10.2196/ipro.4691)

## KEYWORDS

ambulatory blood pressure monitoring; wearable technology; connected health

(This is a conference paper presented at the Connected Health Symposium, Boston, 2015, which was not edited and is only lightly peer-reviewed).

## Multimedia Appendix 1

Extended abstract.

[[PDF File \(Adobe PDF File\), 884KB-Multimedia Appendix 1](#)]

*Edited by G Eysenbach, T Hale; submitted 14.05.15; peer-reviewed by T Tamura; accepted 20.07.15; published 27.10.15*

*Please cite as:*

*Zhang Q, Hu G, Yan X, Chin KY, Strangman GE, Zhao N, Zhang YT*

*Pilot Development of BP-Glass for Unobtrusive Ambulatory Blood Pressure Monitoring*

*iProc 2015;1(1):e8*

*URL: <http://www.iproc.org/2015/1/e8/>*

*doi: [10.2196/ipro.4691](https://doi.org/10.2196/ipro.4691)*

*PMID:*

©Quan Zhang, Gang Hu, Xiangguo Yan, Kong Yien Chin, Gary E. Strangman, Ni Zhao, Yuan-Ting Zhang. Originally published in JMIR Mhealth and Uhealth (<http://www.iproc.org>), 27.10.2015. This is an open-access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/2.0/>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work, first published in JMIR mhealth and uhealth, is properly cited. The complete bibliographic information, a link to the original publication on <http://mhealth.jmir.org/>, as well as this copyright and license information must be included.